


FEATURES

- Detection of low energy, impulsive disturbances (clicks, rub & buzz, etc.)
- Full temporal fine structure of distortion (contrary to Fourier Analysis)
- Reveals short-time disturbances of much lower level than traditional methods
- Applicable to Speakers and Electronics
- Simple interpretation
- Separation from ambient noise
- 3D representation reveals location and physical cause of disturbance
- Uses active Compensation technique for revealing defects
- Measures deviation (impulsive and regular) from a “golden unit”
- Full TRF (standard) features

The TRF-Pro is an enhanced version of the TRF (standard) including all standard features. A new measurement technique is provided for detecting and quantifying low energy, impulsive disturbances (such as clicks and rub & buzz effects). It reveals distortion of much lower level and with a considerably higher temporal resolution (“instantaneous distortion”) than traditional Fourier-analysis, which shows mean values only. Additionally to the common 2D representation (distortion vs. frequency) the distortion can be visualized in a 3D plot. Here the distortion are mapped vs. frequency and vs. one of the measured signals (voice coil displacement, sound pressure, etc.). This is quite useful to localize and identify the cause of the disturbance.

A “golden unit” can be used to define the regular (“good”) behavior for devices under test. Measuring other drivers of the batch will reveal any deviation (impulsive or other) from the defined behavior, which are usually masked by other, dominant effects and which are not detectable using traditional measurement techniques.



Article Number: 1000-910, 1000-911

CONTENTS:

Checking for Disturbances	2
Post-processing	3
Graphical representation	4
Result Windows	4
Limit Values	4
Applications	5


updated August 13, 2012

Checking for Disturbances

Target group

- Design engineers who want to check prototypes for unwanted impulsive disturbances. Furthermore the deviation from defined behavior (“golden unit”) can be checked. This reveals disturbances that are hardly audible or detectable in the system response.
- The Quality Control engineer testing samples from series production. Defective units can be assessed to reveal problems in design, production or material. Long term changes in production of linear and regular nonlinear behavior can be assessed.


Principle



Several distortion components contribute to the measured output signal (sound pressure) of a speaker. As illustrated above, the components contribute at different levels. Linear distortion **L** (linear amplitude and phase response) are much higher than regular distortion **R** caused by motor and suspension nonlinearities. Distortion **D** caused by loudspeaker defects are even lower. After all a certain level of noise **N** is always added to the measured signal. Noise has no correlation with the input signal but can be characterized by a noise floor. Rub & buzz and other effects from defects **D** are usually masked by linear distortion **L** and regular distortion **R**. The measurement principle can be applied not only to speakers but to all devices (DUT) where irregular distortion are masked by regular distortion (electronics, etc.).

Time domain analysis

The parameters of a dedicated loudspeaker model are identified by analyzing the excitation signal $x(t)$ and the output signal $y(t)$. The model is used to reproduce the linear distortion **L** and the regular distortion **R** of the driver. Regular distortion **R** are part of the design and do not contribute to defects **D**. The difference $e(t)$ of both signals holds all information about the distortion **D** and noise **N**.



The error signal $e(t)$ is analyzed by applying the distortion measures explained below.

Measurement without “golden unit”

An adaptive nonlinear model predicts the linear and regular distortion **L+R** in order to subtract them from the output signal **Y**. The residual information are the distortion **D** and noise **N**. Since only one device under test is available, no more information is available to separate noise **N** from distortion **D**.

The error signal $e(t)$ can be analyzed with the distortion measures described below. Note that the full temporal resolution is available.

Measurement with "golden unit"

If one or more "golden units" are available, a model can be learned, that includes the regular (linear and nonlinear) behavior of the units. Using this the regular distortion **L+R** (deterministic part of $e(t)$) can be predicted very accurately and even the typical characteristics of the noise **N** (not deterministic) can be identified.

If a measured DUT does not show any defect the results can be used to update the nonlinear and the noise model (on-the-fly update). Using this learning strategy the algorithm becomes more robust as the model represents more and more the typical regular behavior of the batch.

<h2>Post-processing</h2>	
<h3>Distortion Measures</h3>	
Frequency – time mapping	<p>The TRF uses a logarithmic sine sweep as excitation signal. As at any time instant only a single frequency is excited the measurement time t can be mapped uniquely to the instantaneous excitation frequency f. Hence each distortion measure can either be plotted versus time t or frequency f, as long as the dependency is known. In case of a logarithmic sweep, the linear time scale corresponds to a logarithmic frequency scale. This mapping requires an accurate time delay determination which is done automatically by the TRF.</p> <p>Several distortion measures are provided. They are first calculated in the time domain by post-processing the residual signal $e(t)$ and the measured output signal $y(t)$. The measurement time interval is divided into N subinterval $[t_k, t_{k+1})$, $k=0,1,\dots,N-1$ for which RMS and peak values are calculated. For display the measures are mapped from time domain to frequency domain.</p>
MHD	<p>Mean harmonic distortion</p> $d_{MHD} [k] = \frac{e_{RMS}}{y_{RMS}}$
IHD	<p>Instantaneous harmonic distortion</p> $d_{IHD} (t) = \frac{ e(t) }{y_{RMS}}$
ICHD	<p>Instantaneous crest harmonic distortion</p> $d_{ICHD} (t) = \frac{ e(t) }{e_{RMS}}$
PHD	<p>Peak harmonic distortion</p> $d_{PHD} [k] = \frac{e_{peak}}{y_{RMS}}$
CHD	<p>Crest harmonic distortion</p> $d_{CHD} [k] = \frac{e_{peak}}{e_{RMS}}$

	with $y_{RMS} = \sqrt{\frac{1}{t_{k+1} - t_k} \int_{t_k}^{t_{k+1}} y^2(t) dt}$ $e_{RMS} = \sqrt{\frac{1}{t_{k+1} - t_k} \int_{t_k}^{t_{k+1}} e^2(t) dt}$ $e_{peak} = \max_{t_k \leq t < t_{k+1}} [e(t)]$
DNR	If a “golden unit” is available, for each of the above measures the <u>d</u> istortion to <u>n</u> oise <u>r</u> atio (DNR) can be calculated. This measure indicates deviations (defects) from the behavior of the “golden unit” that are well above the noise floor.
Modes of Operation	
All Distortion	For measurements without “golden unit”. Shows all nonlinear distortion (R+D), i.e. any deviation from linear behavior is considered as distortion. Applications: Electronic circuit testing, long-term variation for regular loudspeaker distortion.
Deviation All Distortion	For measurements with “golden unit”. Shows deviation of nonlinear distortion (R+D) from the distortion of the “golden unit”. A noise model is identified and DNR can be calculated.
Rub & Buzz	For measurements without “golden unit”. Shows impulsive distortion (D). Regular distortion (like motor and suspension distortion) are suppressed. Mainly used for speaker checks.
Deviation Rub & Buzz	For measurements with “golden unit”. Shows deviation of impulsive distortion (D) from the distortion of the “golden unit”. Deviations of regular distortion (like motor and suspension distortion) are not shown. A noise model is identified and DNR can be calculated. Mainly used for speaker checks.

Graphical representation

2D	The selected distortion measure is plotted vs. frequency.
3D (state mapping)	The selected distortion measure is plotted vs. frequency and vs. one of the measured signals (state). The x-axis of the contour plot depicts the frequency, the y-axis shows the measured signal measured and the distortion measure is given in the 3 rd dimension. In case of driver tests the displacement signal is of particular interest (y-axis signal). It can be used to identify the location and thereby the physical cause of the defect.

Result Windows


The following windows are additional to the TRF (standard) result windows.	
Instantaneous Distortion	2D representation of the distortion measures described above versus frequency. Each measure can be scaled in percent or in dB. The distortion measure can be derived from one of the acquired input signals.
Instantaneous Distortion 3D	Shows the selected distortion measure vs. frequency and vs. an acquired signal. The measure can be plotted in percent or in dB. A threshold value can be defined for the 3D coloration of the graph to indicate a defect.
Modeled Response	The measured output signal y(t), the modeled signal y'(t) and the residual error signal e(t) are plotted versus frequency. Over- or under-compensation effects of the error e(t)-signal can be assessed as well as the agreement between modeled and measured output signal.

Limit Values

Parameter	Min	Typ.	Max	Unit
Maximal frequency of nonlinear model			1	kHz
Data acquisition bandwidth		10	43,8	kHz

Applications


High temporal resolution Distortion



With the high temporal resolution the fine structure of distortion can be revealed. Traditional Fourier based methods only provide mean values of the distortion over time. Here the THDN, which is the mean value of the IHD-measure in "All Distortion" mode is plotted for comparison.


Note, that the low frequency oscillations are visible between 10 and 20 Hz. Here the temporal resolution is even higher than the signal period.

Limiting effects in electronic circuits




The instantaneous distortion is also very useful for assessing electronic systems. Here an amplifier produces almost constant total harmonic distortion $THDN \approx -56$ dB over a wide frequency range as shown as bold black line. However, the instantaneous harmonic distortion shown as thin red line reveals narrow peaks up to -40 dB.

The peaks occur in each period of the measured signal as indicated in the time signal (left).



Plotting the instantaneous crest factor of harmonic distortion (ICHHD) versus instantaneous voltage and frequency shows that the distortion are generated at the negative maximum at all frequencies. This is caused by a limiting effect in the amplifier.


Rub and Buzz Detection without "Golden Unit"



The CHD represents the short term peak values of the above ICHHD curve (data reduction).

This measure is well suited for comparing different drivers. It is independent of the amplitude of the distortion but represents as a shape factor the phase information of the higher-order harmonics which is not exploited in traditional measurements.


3D representation



The distortion measure ICHD is plotted vs. displacement and frequency. The ICHD measure is derived from SPL. It is mapped vs. displacement to identify the displacement that corresponds to a particular distortion. In this example a clear rubbing can be detected for voice coil moving outside. The x-axis shows the excitation frequency, the y-axis is the instantaneous displacement of the voice coil and the color

corresponds to the ICHD measure. This reveals valuable information about the location of the R&B defect at 1mm at positive displacement for frequencies below 150 Hz.

Rub and Buzz Detection with "Golden Unit"




Four different golden units have been learned, each sample three times, so the total of learned drivers is 12. A 5th good driver has been tested with the trained model. Here the DNR (blue curve) is around zero.

All of the golden units have some regular (peaky) distortion at 40-100 Hz which is typical for this driver. However, this regular distortion is significantly suppressed by the adaptive compensation technique.


Finally a 6th driver with an hardly audible defect is measured. At low frequencies there is a clear deviation from the trained driver model indicating a defect. The effective DNR is well above 10 dB, exceeding the threshold clearly.

Finding the physical cause with "Golden Unit"



Two defects can be isolated for this driver. At low frequencies at the displacement limits a displacement related defect is visible. This may be caused by staggering. At around 100 Hz a second effect occurs only if the voice coil is inside (limiting by exhausted suspension system).

Variation of nonlinear driver parameters



The TRF-Pro can also be used to monitor the deviation from a given nonlinear behavior. In this example a driver was subjected to a sustained elongation to one displacement limit and was released after 2 minutes. Due to the permanent stretching of the suspension the stiffness nonlinearity has changed and became softer at the rest position. Three measurements were performed, first the original state of the driver before the test. The driver was used as "Golden Unit" to identify a

model of itself. Therefore the distortion noise ratio of the first measurement is about 0 dB (indicating no deviation from the identified model). The second test was performed immediately after releasing the force and shows consequently a high deviation from the original state. After 5 minutes a third experiment was done to check the recovery of the suspension from stress. A clear reduction of the deviation is obvious, however, even after 5 minutes rest the driver did not reach its performance before the test. Valuable clues about the "Memory" effect of the used suspension can be drawn.

Other Applications are parameter stability in long term power tests, temperature dependency or checking the deviation of parameters for a batch of drivers (e.g. from a production line).

Find explanations for symbols at <http://www.klippel.de/know-how/literature.html>

updated August 13, 2012



Klippel GmbH
Mendelssohnallee 30
01309 Dresden, Germany

www.klippel.de
info@klippel.de

TEL: +49-351-251 35 35
FAX: +49-351-251 34 31